

**1. PRODUCT: Keul-o-test Borreliose Complete**

**Description:** Lyme Borreliose IgG + IgM Combo Test Cat # KGST199  
**Quantity:** 202

**2. PHYSICAL INSPECTION: ok**
**3. REAGENT USED:**

	<b>Lot No:</b>	<b>Quantity:</b>	<b>Expiry</b>
a)LYME IgG stick	03089-07	202	2011/04
b)LYME IgM stick	03089-07	202	2011/04
c)Diluent IgG + IgM	25069-02	11 x 11 ml	2012/06

**4. RESULTS:**
**a) LYME IgG**

	<b>Negative serum</b>	<b>SerumNo 140926</b>	<b>Serum No 150956</b>	<b>Serum No 160207</b>	<b>Serum No 160097</b>
<b>Expected (15 min)</b>	-	Shadow	+ / -	+	+
<b>Obtained (15 min)</b>	-	Shadow	+ / -	+	+

**b) LYME IgM**

	<b>Negative serum</b>	<b>SerumNo 150843</b>	<b>Serum No 151219</b>	<b>Serum No 140926</b>	<b>Serum No 150955</b>
<b>Expected (15 min)</b>	-	Shadow	+ / -	+	+
<b>Obtained (15 min)</b>	-	Shadow	+ / -	+	+

**5. RETENTION SAMPLES:**

Retain 5 **Keul-o-test Borreliose Complete** tests as samples per each 6000 tests ordered quantity from each batch (under dessicated conditions) at room temperature until the expiration date.

**6. CONCLUSION:**

Approved      ok  
 Hold  
 Rejected



2009-08-17  
 Date

QM-Manager  
 Günter Keul