

1. PRODUCT: Keul-o-test Borreliose Complete

Description: Lyme Borreliose IgG + IgM Combo Test Cat # KGST199
Quantity: 202

2. PHYSICAL INSPECTION: ok
3. REAGENT USED:

	Lot No:	Quantity:	Expiry
a)LYME IgG stick	12089-01	202	2011/04
b)LYME IgM stick	12089-01	202	2011/04
c)Diluent IgG + IgM	25069-02	10 x 11 ml	2012/06

4. RESULTS:
a) LYME IgG

	Negative serum	SerumNo 140926	Serum No 150956	Serum No 160207	Serum No 160097
Expected (15 min)	-	Shadow	+ / -	+	+
Obtained (15 min)	-	Shadow	+ / -	+	+

b) LYME IgM

	Negative serum	SerumNo 150843	Serum No 151219	Serum No 140926	Serum No 150955
Expected (15 min)	-	Shadow	+ / -	+	+
Obtained (15 min)	-	Shadow	+ / -	+	+

5. RETENTION SAMPLES:

Retain 5 **Keul-o-test Borreliose Complete** tests as samples per each 6000 tests ordered quantity from each batch (under dessicated conditions) at room temperature until the expiration date.

6. CONCLUSION:

Approved ok
 Hold
 Rejected



2009-08-17
 Date

QM-Manager
 Günter Keul