

Product: CD-o-trol Pro<sup>+</sup>  
Lot No.: 050-34L, 050-34H  
Exp. Date: 2010-06-02

Manufactured by: BGT BioGenTechnologies GmbH  
Von-Langen-Weg 10  
D-48565 Steinfurt  
Tel.: 02551/4090 Fax.: 02551/1298

**Product Characteristics:**

This material is a whole blood product and should be similar in appearance to be fresh whole blood after mixing. When stored at 2 to 8°C, unopen vials are stable until the expiration date.  
DO NOT FREEZE.



**WARNING:**

**POTENTIALLY BIOHAZARDOUS MATERIAL.** For *in vitro* diagnostic use. Each human donor/unit used in the preparation of this product has been tested by a FDA licensed method/test and found to be negative or non-reactive for the presence of HBsAg, Anti-HCV, NAT testing for HIV-1, HCV (RNA) and HIV-1/2. Each unit is also negative by a serological test for Syphilis (RPR or STS). Because no test method can offer complete assurance that infectious agents are absent, this material should be handled as human blood. When handling or disposing of vials follow precautions for patient specimens as specified in the OSHA Bloodborne Pathogen Rule (OSHA 29 CFR Part 1910.1030) or other equivalent biosafety procedures.

2. Lot uniformity studies verified product specifications were met and maintained in vial-to-vial testing.
3. Fill volumes were verified.
4. Microbiology testing meets specification. These lots have been tested as per quality control procedures and found free of viable pathogenic or hemolytic microorganisms and fungi. This material is non-sterile.

This product has been manufactured according to all applicable FDA Good Manufacturing Practice regulations. These lots meet Manufacturing and Quality Assurance release specifications and are approved for release to Shipping/Marketing.



QM-Manager  
Günter Keul

2010-04-15  
Date