

**1. PRODUCT: Keul-o-test Clostridium Difficile Complete**

**Description:** Clostridium Difficile A + Clostridium Difficile B Combo Test Cat # KGST203  
**Quantity:** 404

**2. PHYSICAL INSPECTION: ok**
**3. REAGENT USED:**

	<b>Lot No:</b>	<b>Quantity:</b>	<b>Expiry</b>
a) Toxin A stick Clostridium Difficile A	07079-02	404	2011/03
b) Toxin B stick Clostridium Difficile B	07079-02	404	2011/03
c) Extraction solution	11069-08	404 x 2 ml	2012/06

**4. RESULTS:**
**a) TOXIN A (Clostridium Difficile A)**

	<b>Negative</b>	<b>Standard No1</b>	<b>Standard No2</b>	<b>Standard No3</b>	<b>Standard No4</b>
<b>Expected after 15 min</b>	-	-	+ or -	+	+
<b>Obtained after 15 min</b>	-	-	-	+	+

**b) TOXIN B (Clostridium Difficile B)**

	<b>Negative</b>	<b>Standard No1</b>	<b>Standard No2</b>	<b>Standard No3</b>	<b>Standard No4</b>
<b>Expected after 15 min</b>	-	-	+ or -	+	+
<b>Obtained after 15 min</b>	-	-	-	+	+

**5. RETENTION SAMPLES:**

Retain 6 **Keul-o-test Clostridium Difficile Complete** tests as samples per each 6000 tests ordered quantity from each batch (under dessicated conditions) at room temperature until the expiration date.

**6. CONCLUSION:**

Approved      ok  
 Hold  
 Rejected



2009-06-27  
 Date

QM-Manager  
 Günter Keul