

(Revised 10/18/2002)

Product Name:	Drog-o-test Multiline 5 Testkarte	KGST129	Max Reading Time	5'
Product Format:	Cassette		Specimen:	Urine
Lot Number:	20100801		Exp Date:	2013-01
Quantity:				

Raw materials:

1. All reagents and raw materials used in the manufacture of the one step tests are qualified for use by many different parameters.
 2. Only qualified materials are released to manufacturing for product preparation. 3. All the reagents are uniformly dissolved in the solution during the production process.

Visual Result Description:

Positive: ___1___ colored lines visible in the result window within the maximum reading time.

Negative: ___2___ colored lines visible in the result window within the maximum reading time.

Invalid: No lines visible in the result window within the maximum reading time.

Storage: Product must be stored at room temperature avoid direct sunlight.

Functional Quality Inspection:

The functional ability of the product is tested by randomly chosen samples from the finished products. Follow the procedure specified in the insert.

Detailed visual test results

ITEM	Sensitivity	negative control								positive control							
		1		2		3		4		5		6		7		8	
		C	T	C	T	C	T	C	T	C	T	C	T	C	T	C	T
AMP	1000 ng/ml	18"	25"	20"	27"	18"	25"	20"	27"	20"	-	22"	-	20"	-	22"	-
OPI	300 ng/ml	26"	14"	28"	16"	26"	14"	28"	16"	26"	-	28"	-	26"	-	28"	-
MET	1000 ng/ml	19"	14"	21"	16"	19"	14"	21"	16"	21"	-	23"	-	21"	-	23"	-
THC	50 ng/ml	37"	34"	39"	36"	37"	34"	39"	36"	35"	-	37"	-	35"	-	37"	-
COC	300 ng/ml	14"	16"	16"	18"	14"	16"	16"	18"	17"	-	19"	-	17"	-	19"	-

Conclusion: Pass ___x___

Not Pass _____

This batch of the product is assigned to the following lot number and expiration date.



 QM-Manager
 Günter Keul

2010-09-05

Date

09SM02005