

(Revised 10/18/2002)

Product Name:	Drog-o-test Multi Drogen 10	GKDOA125 - GKDOA135 - GKDOA145 - GKDOA155 - GKDOA165 - GKDOA175 - GKDOA185 - GKDOA195 - GKDOA1105	Max Reading Time	5'
Product Format:	Cassette		Specimen:	Urine
Lot Number:	201103116		Exp Date:	2013-08
Quantity:				

Raw materials:

1. All reagents and raw materials used in the manufacture of the one step tests are qualified for use by many different parameters.
 2. Only qualified materials are released to manufacturing for product preparation. 3. All the reagents are uniformly dissolved in the solution during the production process.

Visual Result Description:

Positive: ___1___ colored lines visible in the result window within the maximum reading time.
 Negative: ___2___ colored lines visible in the result window within the maximum reading time.
 Invalid: No lines visible in the result window within the maximum reading time.

Storage: Product must be stored at room temperature avoid direct sunlight.

Functional Quality Inspection:

The functional ability of the product is tested by randomly chosen samples from the finished products. Follow the procedure specified in the insert.

Detailed visual test results

ITEM	Sensitivity	negative control								positive control							
		1		2		3		4		5		6		7		8	
		C	T	C	T	C	T	C	T	C	T	C	T	C	T	C	T
AMP	1000ng/ml	23"	12"	25"	14"	23"	12"	25"	14"	18"	-	20"	-	18"	-	20"	-
COC	300ng/ml	22"	13"	24"	15"	22"	13"	24"	15"	15"	-	17"	-	15"	-	17"	-
MET	1000ng/ml	38"	27"	40"	29"	38"	27"	40"	29"	20"	-	22"	-	20"	-	22"	-
OPI	300ng/ml	18"	8"	20"	10"	18"	8"	20"	10"	21"	-	23"	-	21"	-	23"	-
THC	50ng/ml	27"	18"	29"	20"	27"	18"	29"	20"	32"	-	34"	-	32"	-	34"	-
BAR	300ng/ml	20"	12"	22"	14"	20"	12"	22"	14"	22"	-	24"	-	22"	-	24"	-
BZD	200ng/ml	23"	18"	25"	20"	23"	18"	25"	20"	23"	-	25"	-	23"	-	25"	-
MTD	300ng/ml	30"	16"	32"	18"	30"	16"	32"	18"	18"	-	20"	-	18"	-	20"	-
PCP	25ng/ml	22"	42"	24"	44"	22"	42"	24"	44"	31"	-	33"	-	31"	-	33"	-
TCA	1000ng/ml	19"	17"	21"	19"	19"	17"	21"	19"	26"	-	28"	-	26"	-	28"	-

Conclusion: Pass ___x___ Not Pass _____

This batch of the product is assigned to the following lot number and expiration date.



QM-Manager
Günter Keul

2011-04-12
Date

09SM02005