



BGT
BioGenTechnologies
GmbH
 Von-Langen-Weg 10
 48565 Steinfurt
 e-mail: info@biogentechnologies.de
 web: www.biogentechnologies.de

Certificate of Analysis

(Revised 10/18/2002)

Product Name:	FemLab	Manufacture Date:	211/28/08
Product Parameter:	PH (X)	Gardnerella (X)	NIT (X)
	BLO (X)	PRO (X)	LEU (X)
Lot Number:	20080901	Exp. Date:	2010-02

Testing Date: 11/28/08	Total Qualities tested: 12	Control Manufacturer: Bayer
Quantities tested for positive control: 6	Quantities tested for negative control: 6	Instrument:

Testing Result for Positive Control:

	1	2	3	4	5	6
PH	6.0	6.0	6.0	6.0	6.0	6.0
Gardnerella	Positive	Positive	Positive	Positive	Positive	Positive
NIT	Positive	Positive	Positive	Positive	Positive	Positive
BLO	Positive	Positive	Positive	Positive	Positive	Positive
PRO	Positive	Positive	Positive	Positive	Positive	Positive
LEU	Positive	Positive	Positive	Positive	Positive	Positive
REMARK	-	-	-	-	-	-

Testing Result for Negative Control:

	1	2	3	4	5	6
PH	4.0	4.0	4.0	4.0	4.0	4.0
Gardnerella	Negative	Negative	Negative	Negative	Negative	Negative
NIT	Negative	Negative	Negative	Negative	Negative	Negative
BLO	Negative	Negative	Negative	Negative	Negative	Negative
PRO	Negative	Negative	Negative	Negative	Negative	Negative
LEU	Negative	Negative	Negative	Negative	Negative	Negative
REMARK	-	-	-	-	-	-

Positive percentage _____ 100% _____

Negative percentage _____ 100% _____

Conclusion (X) The products **Pass** The QC tests.
 () The Products **Do Not Pass** the QC tests.

QM-Manager
 Günter Keul

2008-03-30

Date

09SM02005