



BGT BioGenTechnologies GmbH
 Von-Langen-Weg 10
 48565 Steinfurt
 e-mail: info@biogentechnologies.de
 web: www.biogentechnologies.de

**Certificate of
 Analysis**

(Revised 10/18/2002)

Product Name	Keul-o-test Glucose/Keton	Manufacture Date: 12/10/09			
Product Parameter:	GLU (x)	BIL ()	KET (x)	SP ()	BLO ()
	pH ()	PRO ()	URO ()	NIT (x)	LEU ()
Lot Number:	20091104			Exp. Date:	2011-04

Testing Date: 12/10/09	Total Quantities tested:12	Control Manufacturer: Bayer
Quantities tested for Positive Control:6	Quantities tested for negative control:6	Instrument:

Testing Result for Positive Control:

	1	2	3	4	5	6
GLU	Trace	Trace	Trace	Trace	Trace	Trace
BIL	-	-	-	-	-	-
KET	+++	+++	+++	+++	+++	+++
SP	-	-	-	-	-	-
BLO	-	-	-	-	-	-
PH	-	-	-	-	-	-
PRO	-	-	-	-	-	-
URO	-	-	-	-	-	-
NIT	-	-	-	-	-	-
LEU	-	-	-	-	-	-
REMARK						

Testing Result for Negative Control:

	1	2	3	4	5	6
GLU	Negative	Negative	Negative	Negative	Negative	Negative
BIL	-	-	-	-	-	-
KET	Negative	Negative	Negative	Negative	Negative	Negative
SP	-	-	-	-	-	-
BLO	-	-	-	-	-	-
PH	-	-	-	-	-	-
PRO	-	-	-	-	-	-
URO	-	-	-	-	-	-
NIT	-	-	-	-	-	-
LEU	-	-	-	-	-	-
REMARK						

Positive percentage 100% Negative Percentage 100%

Conclusion: (x) The products **Pass** the QC tests.
 () The products **Do Not Pass** the QC tests.

QM-Manager
 Günter Keul

2009-12-13
 Date

09SM02005