

The Quality Control Department at BGT BioGenTechnologies GmbH certifies that this product has been tested and meets all performance specifications and release requirements.

Date of manufacture:	2009-02
Kit description:	Keul-o-trol H Richtigkeitskontrolle Human Level 2 Human serum normal
Catalogue No.:	KG1530
Batch /Lot No.:	566UN
Expiry date:	212-06

Analyte	Target	Range	Result	Units
Albumin	39.8	N/A	41.3	g/l
Total Protein	59.3	N/A	62.3	g/l

Human source material from which this product derived has been tested at donor level for HIV1 & HIV2, Syphilis, HbsAG and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved by:



Quality control representative
Günter Keul

2009-05-13
Date