

PRODUCT: VET 510 Vet-o-test Leishmania
 Lot No: AI110901
 EXP: 2011.01.05

DATE OF ISSUE: 2009.07.17
 EFFICACY TEST DATE: 2009.07.10
 PACKAGING MATERIALS TEST: 2009.07.14

EFFICACY TEST

Replicate No	Standard panel						Result
	1		2		3		
Negative buffer	-		-		-		Passed
Low titer	+		+		+		Passed
Medium titer	+		+		+		Passed
High titer	+		+		+		Passed

Specimen No	Spiking (3 times)					Result
	1	2	3	4	5	
Medium titer	+	+	+	+	+	Passed

Specimen No	Negative specimen										Result	
	1	2	3	4	5	6	7	8	9	10		
Serum, Plasma, Whole blood	-	-	-	-	-	-	-	-	-	-	-	Passed

FINISHED PACKAGING MATERIALS TEST: Conformed

STORAGE CONDITION: Store at room temperature. Do not refrigerate or freeze.

FINAL CONTROL PROCEDURE: Conformed

ANALYSIT CERTIFICATION: Department of Quality Control

CERTIFICATE APPROVAL: Department of Quality Control



QM-Manager
 Günter Keul