

Product name: Keul-o-test PSA K
DIMDI Reg.-Nr.: DE/CA22/1116-110-IVD

Catalogue number: KGST183

Quantity: 20 tests in a case

Batch number: PLPSAC1005

Expiration date: 2013-06-15

Specimen: Plasma/Serum or Whole blood

Shelf life: 18 month

Storage & Shipping condition: 1-30°C (34-86°F) in a dry place, Do not freeze.

Sensitivity: 4 ng/mL of total PSA

Quality Control results

Test Category	Inspection Criteria	Test Result
Specificity Test	Negative	Negative
Sensitivity Test	Positive	Positive
Blood Separation	Clear background	Passed



QM-Manager
Günter Keul

2011-12-26
Date