

1. PRODUCT: Keul-o-test R.S.V.

Description: R.S.V. Test Cat # KGST200
Quantity: 202

2. PHYSICAL INSPECTION: ok
3. REAGENT USED:

	Lot No:	Quantity:	Expiry
a) device	31039-01	202	2010/11
b) Extraction solution	04128-13	21 x 11 ml	2011/12

4. RESULTS:

Standards	Extraction solution	1/800	1/400	1/200	1/100
Expected after 10 min	-	-	+ / -	+f	+
Obtained after 10 min	-	-	+ / -	+f	+

5. RETENTION SAMPLES:

Retain 6 **Keul-o-test R.S.V.** tests as samples per each 6000 tests ordered quantity from each batch (under dessicated conditions) at room temperature until the expiration date.

6. CONCLUSION:

Approved ok
 Hold
 Rejected



QM-Manager
 Günter Keul

2009-04-03
 Date