

(Revised 10/18/2002)

Product Name:	StrepA	KGST109	Max Reading Time	10'
	DIMDI Reg.-Nr.: DE/CA22/1116-137-IVD			
Product Format:	Dipstrip ()	Cassette (x)	Midstream ()	Other ()
Specimen:	Urine ()	Serum ()	Whole Blood ()	Other (x)
Negative Control Lot:	11010501-8		Sensitivity:	N/A
Positive Control Lot:	08020101		Quantity:	

Raw materials:

1. All reagents and raw materials used in the manufacture of the one step tests are qualified for use by many different parameters.
 2. Only qualified materials are released to manufacturing for product preparation. 3. All the reagents are uniformly dissolved in the solution during the production process.

Visual Result Description:

 Positive: 2 colored lines visible in the result window within the maximum reading time.

 Negative: 1 colored lines visible in the result window within the maximum reading time.

Invalid: No lines visible in the result window within the maximum reading time.

Storage: Product must be stored at room temperature avoid direct sunlight.

Functional Quality Inspection:

The functional ability of the product is tested by randomly chosen samples from the finished products. Follow the procedure specified in the insert.

a. Detailed visual test results using negative control:

Sample No.	Negative Control or Specimen	"Test Line" Observed at	"Control Line" Observed at	Background Clearance Observed at
1	-HUMAN CELL	- WITHIN 10'	+ AT 37"	3'
2	-HUMAN CELL	- WITHIN 10'	+ AT 39"	3'
3	-HUMAN CELL	- WITHIN 10'	+ AT 37"	3'
4	-HUMAN CELL	- WITHIN 10'	+ AT 39"	3'

b. Detailed visual test results using positive control:

Sample No.	Positive Control or Specimen	"Test Line" Observed at	"Control Line" Observed at	Background Clearance Observed at
5	+HUMAN CELL	+ AT 1'33"	+ AT 43"	3'
6	+HUMAN CELL	+ AT 1'35"	+ AT 45"	3'
7	+HUMAN CELL	+ AT 1'33"	+ AT 43"	3'
8	+HUMAN CELL	+ AT 1'35"	+ AT 45"	3'

 Conclusion: Pass x

Not Pass _____

This batch of the product is assigned to the following lot number and expiration date.

Lot Number: 201104086
Exp Date: 2013-09


 QM-Manager
 Günter Keul

2011-05-22

Date

09SM02005