

1. PRODUCT: Keul-o-test Tetanus

Description: Anti-tetanus anatoxine antibody rapid test Cat # KGST204
Quantity: 202

2. PHYSICAL INSPECTION: ok
3. REAGENT USED:

	Lot No:	Quantity:	Expiry
a) device	28071-04	202	2013/03
b) Extraction solution	11071-04	11 x 5 ml	2014/07

4. RESULTS:

Standards	Negative pool	W.H.O. serum control	Standard No 1	Positive pool
Expected (20 minutes)	-	+ or -	+	+
Obtained (10 minutes)	-	-	+	+

5. RETENTION SAMPLES:

Retain 10 **Keul-o-test Tetanus** tests as samples per each single order from each batch (under dessicated conditions) at room temperature until the expiration date.

6. CONCLUSION:

Approved ok
 Hold
 Rejected



QM-Manager
 Günter Keul

2011-08-06
 Date