



QMLab GmbH

Von-Langen-Weg 10
48565 Steinfurt
Tel.: 02551-149913
Fax.:02551-149914

Certificate of Analysis

RE-TYPED FROM THE ORIGINAL CERTIFICATE

Date: 2009-01-16

Product: Troponin I Device

PRODUCT CODE/ ART.-NO.	LOT NUMBER	EXPIRATION DATE	QUANTITY	PHYSICAL APPEARANCE
QML124	4D097-8125	2011-01	5005	conforms

	Solution	Amount	Result		Pass Y/N
			positive	negative	
Negative	0 ng/ml Trop I /ml	3	0	3	y
low Control	1 ng Trop I (Hytest)/ml	3	3	0	y
Positive Control	20 ng Trop I (Hytest)/ml	3	3	0	y
physical appearance					y

PRODUCT MANUFACTURING

The above listed product has been manufactured in accordance with the relevant standards (DIN EN ISO 13485:2003). No radioactive material of any kind is utilized in the product, or in the manufacturing of the product.

The above listed product does not contain any virus, reactive by-product of the same, or metabolic by-product of hepatitis A, B, C, or D or of the antibodies for the human acquired immune deficiency syndrome (AIDS).

If this product is shipped as bulk, QMLab GmbH assumes no responsibility for the packing and label claims.

QM-Management
Günter Keul