



BGT
BioGenTechnologies
GmbH

Von-Langen-Weg 10
48565 Steinfurt
e-mail: info@biogentechnologies.de
web: www.biogentechnologies.de

Certificate of Analysis

(Revised 10/18/2002)

Product Name	Ur-o-test 10S				Manufacture Date:	06/03/08
Product Parameter:	GLU (x)	BIL (x)	KET (x)	SP (x)	BLO (x)	
	pH (x)	PRO (x)	URO (x)	NIT (x)	LEU (x)	
Lot Number:	20080205				Exp. Date:	2009-07

Testing Date: 03/28/08	Total Quantities tested:12	Control Manufacturer:Bayer
Quantities tested for Positive Control:6	Quantities tested for negative control:6	Instrument:

Testing Result for Positive Control:

	1	2	3	4	5	6
GLU	+	+	+	+	+	+
BIL	+	+	+	+	+	+
KET	+++	+++	+++	+++	+++	+++
SP	≤ 1.005	≤ 1.005	≤ 1.005	≤ 1.005	≤ 1.005	≤ 1.005
BLO	+++	+++	+++	+++	+++	+++
PH	8.0	8.0	8.0	8.0	8.0	8.0
PRO	+	+	+	+	+	+
URO	2.0EU/dl	2.0EU/dl	2.0EU/dl	2.0EU/dl	2.0EU/dl	2.0EU/dl
NIT	Positive	Positive	Positive	Positive	Positive	Positive
LEU	+	+	+	+	+	+
REMARK	-	-	-	-	-	-

Testing Result for Negative Control:

	1	2	3	4	5	6
GLU	Negative	Negative	Negative	Negative	Negative	Negative
BIL	Negative	Negative	Negative	Negative	Negative	Negative
KET	Negative	Negative	Negative	Negative	Negative	Negative
SP	1.020	1.020	1.020	1.020	1.020	1.020
BLO	Negative	Negative	Negative	Negative	Negative	Negative
PH	6.0	6.0	6.0	6.0	6.0	6.0
PRO	Negative	Negative	Negative	Negative	Negative	Negative
URO	0.2EU/dl	0.2EU/dl	0.2EU/dl	0.2EU/dl	0.2EU/dl	0.2EU/dl
NIT	Negative	Negative	Negative	Negative	Negative	Negative
LEU	Negative	Negative	Negative	Negative	Negative	Negative
REMARK	-	-	-	-	-	-

Positive percentage _____ 100% _____ Negative Percentage _____ 100% _____

Conclusion: (x) The products **Pass** the QC tests.
() The products **Do Not Pass** the QC tests.

QM-Manager
Günter Keul

2008-03-08
Date

09SM02005